

CERTIFICATE



**Certificate to Letter of assessment  
VN635 171792**

in accordance with enactment GZ 2020-0.247.451 as well as according to "Test principle for Corona SARS-Cov-2 pandemic respirators Rev. 0 dated 20.03.2020" for placing on the market in Austria

---

Based on the documents submitted and the examinations carried out this is to certify, that the Corona Virus Pandemic Respirator (CPA)

**"Atemschutzmaske FFP2 / N95 / KN95"**

*(photo documentation see reverse side)*

corresponds to the requirements of the shortened assessment process in accordance with enactment GZ 2020-0.247.451 as well as according to "Test principle\*) for Corona SARS-Cov-2 pandemic respirators Rev. 0 dated 20.03.2020" for placing on the market in Austria based on EN 149:2001+A1:2009, the following can be stated:

**The Corona Virus Pandemic Respirator (CPA) passed the test with a positive result <sup>note)</sup>**

<sup>note)</sup> The fit of the mask depends on the face shape of the wearer, so a fit check must be carried out before use. Appropriate information must be enclosed with the mask.

This new PPE will be assigned to **category III**

This letter of assessment applies only to Corona Virus Pandemic Respirator (CPA) to medical professionals for the duration of the current health threat and / or until the GZ 2020-0.247.451 decree expires.



Ing. Judith Pointner  
Head of Certification Body

OETI - Institute for Ecology, Technology and Innovation; Notified Body No. NB 0534  
Vienna, 31.07.2020

\*) non-harmonized standard